GERT CATE

Certificate-Register-No.: Z - 09 - 107 - MP - E

Berlin Cert GmbH

Institute for Testing and Certification of Medical Devices associated with the Technical University Berlin

certifies, that the following product of the company

AAT Alber Antriebstechnik GmbH

Ehestetter Weg 11, D-72458 Albstadt

fulfills the requirements of the named standards

Product: Stair-climbing device S-max D1351

Requirements: DIN EN 12182:1999

Technical aids for disabled persons - General requirements and test methods

ISO 7176-23:2002

Wheelchairs - Part 23: Requirements and test methods for attendant-operated stair-climbing devices

The conducted tests are documented in the test protocols P-09-107-MP-PA041 and P-09-107-MP-PA024.

This certificate is not a confirmation, that the product complies with the German medical device law or the directive 93/42/EEC

Issued on: 2009-04-24